Ambulatory 24-h reflux monitoring: a comparison between glass-, antimonyand ISFET electrodes

Published: 22-07-2008 Last updated: 06-05-2024

To compare oesophageal acid exposure time measured with glass, ISFET and antimony electrodes in patients with reflux symptoms.

Ethical review	Approved WMO
Status	Recruitment stopped
Health condition type	Gastrointestinal conditions NEC
Study type	Observational invasive

Summary

ID

NL-OMON32307

Source ToetsingOnline

Brief title pH electrodes reflux monitoring

Condition

• Gastrointestinal conditions NEC

Synonym GORD, heartburn

Research involving Human

Sponsors and support

Primary sponsor: Universitair Medisch Centrum Utrecht Source(s) of monetary or material Support: Janssen-Cilag

Intervention

Keyword: GORD, pH electrodes, pH monitoring

Outcome measures

Primary outcome

Acid exposure time (percentage of time with a pH below 4 at 5 cm above the

upper border of the lower oesophageal sphincter)

Secondary outcome

The numbers reflux episodes detected by impedance monitoring subdivided into

acid, weakly acidic and weakly alkaline reflux episodes using pH measurement.

Study description

Background summary

The acid exposure time is an important parameter that is assessed during 24-h pH-impedance monitoring in patients with gastro-oesophageal reflux disease (GORD). It is expressed as the percentage of time with a pH below the threshold of 4 in the oesophagus at 5 cm above the upper margin of the lower oesophageal sphincter (LOS). Several types of pH electrodes (glass, antimony and ISFET electrodes) are currently commercially available. It is not known whether these different types of pH electrodes will result in different acid exposure times in patients with symptoms of reflux disease.

Study objective

To compare oesophageal acid exposure time measured with glass, ISFET and antimony electrodes in patients with reflux symptoms.

Study design

We will simultaneously record oesophageal pH with three different pH electrodes in each patient during an ambulatory 24h pH-impedance monitoring.

Study burden and risks

After placement, the three catheters cause no additional risk or inconvenience compared to the regular measurement with one catheter. The total diameter of the three catheters is comparable with the diameter of the catheters used during ambulatory pH-manometry in patients with non-cardiac chest pain.

Contacts

Public Universitair Medisch Centrum Utrecht

Heidelberglaan 100 3584CX Utrecht Nederland **Scientific** Universitair Medisch Centrum Utrecht

Heidelberglaan 100 3584CX Utrecht Nederland

Trial sites

Listed location countries

Netherlands

Eligibility criteria

Age

Adults (18-64 years) Elderly (65 years and older)

Inclusion criteria

Reflux symptoms Age > 18y Informed consent

Exclusion criteria

Age < 18y Absence of Informed consent

Study design

Design

Study type: Observational invasive		
Masking:	Open (masking not used)	
Control:	Uncontrolled	
Primary purpose:	Diagnostic	

Recruitment

NL	
Recruitment status:	Recruitment stopped
Start date (anticipated):	12-08-2008
Enrollment:	23
Туре:	Actual

Ethics review

Approved WMO	
Date:	22-07-2008
Application type:	First submission
Review commission:	METC Universitair Medisch Centrum Utrecht (Utrecht)

Study registrations

Followed up by the following (possibly more current) registration

No registrations found.

4 - Ambulatory 24-h reflux monitoring: a comparison between glass-, antimony- and IS ... 21-06-2025

Other (possibly less up-to-date) registrations in this register

No registrations found.

In other registers

Register

ССМО

ID NL23452.041.08