

# The reliability of a new guidance tool voor accurate cuppositioning in total hip arthroplasty

Published: 07-05-2013

Last updated: 24-04-2024

In deze studie zal de nauwkeurigheid van peroperatieve meting van de cuppositie door een nieuw ontworpen, onlangs gepatenteerd, meetinstrument worden onderzocht. In de toekomst kan, indien het meetinstrument betrouwbaar blijkt, peroperatief de...

<b>Ethical review</b>	Approved WMO
<b>Status</b>	Pending
<b>Health condition type</b>	Other condition
<b>Study type</b>	Interventional

## Summary

### ID

NL-OMON39661

### Source

ToetsingOnline

### Brief title

Guidewire trial

### Condition

- Other condition
- Joint disorders

### Synonym

arthrosis, osteoarthritis

### Health condition

gewrichtskraakbeen

### Research involving

Human

## Sponsors and support

**Primary sponsor:** Isala Klinieken

**Source(s) of monetary or material Support:** HipSecure b.v.

## Intervention

**Keyword:** accuracy, cup, guidancetool, total hip arthroplasty

## Outcome measures

### Primary outcome

the difference of perioperative measured inclination and anteversion (by using the device) compared to the values calculated on CT.

### Secondary outcome

operating time, complications, such as: infection and pain at the entrancepoint of the markingpoint.

## Study description

### Background summary

Osteoarthritis of the hip is one of the most common disease in Holland. About 257.400 people in the Netherlands are suffering from this disease. In the Isala Clinics about 700 primary total hip arthroplasties are performed each year. Because osteoarthritis is age related the demand for this procedure will increase. Joint replacing surgery is now world wide one of the most performed procedures. For this reason the occurrence of complications should be kept to a minimum.

The most important complications are luxation, infection and loosening of components. Malpositioning of the cup is highly related to luxation and loosening of components.

The orientation of the acetabular component is defined in anteversion and inclination. Anteversion is the angle used to define the version of the opening of the cup to the front of the pelvis. Inclination is the angle which the cup forms with a horizontal line. Lewinnek defined a \*safe zone\* in which the cup would have an optimal position in the pelvis. This safe zone is defined as: 15

+/- 10 degrees anteversion and 45 +/- 10 degrees inclination. In this safe zone there is an optimal correlation between \*range of motion\* and stability of the prosthesis. Correct placement of the cup is usually difficult because of the absence of reliable intra-operative "landmarks". For this reason, about 50% of the cups are placed in a different position than the surgeon estimated while placing the cup during the procedure. It is clear that a method for more accurate cuppositioning is highly needed.

## **Study objective**

In deze studie zal de nauwkeurigheid van peroperatieve meting van de cuppositie door een nieuw ontworpen, onlangs gepatenteerd, meetinstrument worden onderzocht. In de toekomst kan, indien het meetinstrument betrouwbaar blijkt, peroperatief de cuppositie op geleide van de metingen van het instrument worden aangepast en dus het postoperatieve complicatierisico verlagen.

In this study the accuracy of a new device will be tested for placement of the acetabular component in total hip arthroplasty. In the future this instrument can contribute to more accurate placement of the acetabular component, so the incidence of complications can be decreased.

## **Study design**

prospective cohort study.

## **Intervention**

placement of a screw in the pelvis, and increase of operationtime of approximately 10 minutes.

## **Study burden and risks**

during the operation a screw is inserted in the bone. The screw is about 2 mm in diameter and is placed in the pelvis (Spina Iliaca Anterior Superior). . The depth of these pins in the bone is about 1 to 2.0 cm. the risk of this intervention can be expressed in postoperative pain and infection. A few days after the procedure a CT-scan will be made. This exposes the patient to some degree of radiation.

The benefit of participation is that the precise orientation of the acetabular component is known (CT-scan). This is of great importance for the prognosis of the implant. If malpositioning has occurred we can instruct the patient so that the risk of complications can be reduced.

## Contacts

### Public

Isala Klinieken

Admiraal De Ruijterweg 395-2hg  
amsterdam 1055mc  
NL

### Scientific

Isala Klinieken

Admiraal De Ruijterweg 395-2hg  
amsterdam 1055mc  
NL

## Trial sites

### Listed location countries

Netherlands

## Eligibility criteria

### Age

Adults (18-64 years)

Elderly (65 years and older)

### Inclusion criteria

osteoarthritis of the hip

### Exclusion criteria

obesity

hip dysplasia

traumatic hip injury

patients with increased risk for infection

## Study design

### Design

**Study type:** Interventional

Masking: Open (masking not used)

Control: Uncontrolled

Primary purpose: Prevention

### Recruitment

NL

Recruitment status: Pending

Start date (anticipated): 01-04-2013

Enrollment: 25

Type: Anticipated

### Medical products/devices used

Generic name: insertion of a screw in the pelvis; on the screw a guidance tool will be attached

Registration: Yes - CE intended use

## Ethics review

Approved WMO

Date: 13-05-2013

Application type: First submission

Review commission: METC Isala Klinieken (Zwolle)

## Study registrations

### Followed up by the following (possibly more current) registration

No registrations found.

## Other (possibly less up-to-date) registrations in this register

No registrations found.

## In other registers

### Register

CCMO

### ID

NL42724.075.12