

Sensor for Measuring the blood glucose level as input for the ARtificial pancreas: a sensor validation Trial.

Published: 28-09-2020

Last updated: 09-04-2024

The main objective is to determine the accuracy of the Guardian Sensor 3 in combination with the transmitter developed by Inreda Diabetic BV. Secondary objectives are to assess differences in sensor performance parameters between the Guardian Sensor...

Ethical review	Approved WMO
Status	Recruitment stopped
Health condition type	Glucose metabolism disorders (incl diabetes mellitus)
Study type	Interventional

Summary

ID

NL-OMON49011

Source

ToetsingOnline

Brief title

SMART 1

Condition

- Glucose metabolism disorders (incl diabetes mellitus)

Synonym

Diabetes type 1

Research involving

Human

Sponsors and support

Primary sponsor: Inreda Diabetic B.V.

Source(s) of monetary or material Support: Bedrijf: Inreda Diabetic

Intervention

Keyword: Diabetes, Glucose, Sensor, Validation

Outcome measures

Primary outcome

Main parameter to express glucose accuracy is the MARD value (mean absolute relative difference between self-monitoring of blood glucose (SMBG) measurements and Guardian Sensor 3 measurements).

Secondary outcome

Secondary accuracy parameters are:

- * Overall percentage of measured glucose values that fall within ± 0.83 mmol/l of a SMBG measurement at a glucose level of < 5.55 mmol/l.
- * Overall percentage of measured glucose values that fall within 15% of a SMBG measurement at a glucose level of >5.55 mmol/l.
- * Overall percentage of measured glucose values that fall in zone A and B of the Clarke Error Grid.
- * PARD value of the Guardian Sensor 3 in combination with the Enlite Sensor.

AP-related outcome parameters are:

- * Life-time of the sensors;
- * Percentage of time that the sensors are available as input for the algorithm.

Other study parameters are demographic characteristics, weight, length, HbA1c and amount of administered insulin during usual diabetes treatment.

Study description

Background summary

Inreda Diabetic BV. (Goor, The Netherlands) developed a bi-hormonal reactive close loop system to automate control glucose regulation in patients with diabetes mellitus type 1. This artificial pancreas (AP) consists of two subcutaneous glycosensor enzyme sensors with transmitters to measure continuously the blood glucose values and consists of an algorithm to determine the amount of insulin or glucagon to be administered. In the current CE-marked AP, glucose values are measured by Enlite Sensor (Medtronic) in combination with Inreda Diabetic transmitters. In the future, Medtronic will replace the Enlite Sensor by the Guardian Sensor 3 (Medtronic).

Study objective

The main objective is to determine the accuracy of the Guardian Sensor 3 in combination with the transmitter developed by Inreda Diabetic BV. Secondary objectives are to assess differences in sensor performance parameters between the Guardian Sensor 3 and Enlite Sensor both in combination with Inreda transmitter; to explore AP-related outcomes for both the Guardian Sensor 3 and Enlite Sensor.

Study design

This study is a single center, prospective, observational, sensor validation trial.

Intervention

not applicable

Study burden and risks

There are no major risks associated with this study since the patients will use their usual diabetes therapy during the study. The most prominent risk is that patients may develop irritation and redness caused by the plaster to attach the sensors and possibly hematoma or pain from the SMBGs/sensors. The burden of the study is also relatively low. Patients have to visit the CRC only twice and are only asked to perform and register 7 SMBG measurements per day. There are no individual benefits for participating in this study.

Contacts

Public

Inreda Diabetic B.V.

Klavermaten 65-5

Goor 7472 DD

NL

Scientific

Inreda Diabetic B.V.

Klavermaten 65-5

Goor 7472 DD

NL

Trial sites

Listed location countries

Netherlands

Eligibility criteria

Age

Adults (18-64 years)

Elderly (65 years and older)

Inclusion criteria

- * Diagnosed with diabetes mellitus type 1;
- * Age between 18 and 75 years;
- * Willing and able to sign informed consent.

Exclusion criteria

- * BMI > 35 kg/m²;
- * Skin condition prohibiting needle insertion;
- * Known or suspected problem related to enzyme based glucose sensor usage;
- * Use of acetaminophen during the study period;

* Patients who go swimming or visit a sauna during the study period.

Study design

Design

Study type: Interventional

Masking: Open (masking not used)

Control: Uncontrolled

Primary purpose: Other

Recruitment

NL

Recruitment status: Recruitment stopped

Start date (anticipated): 11-11-2020

Enrollment: 10

Type: Actual

Medical products/devices used

Generic name: transmitter of the Artificial Pancreas

Registration: Yes - CE intended use

Ethics review

Approved WMO

Date: 28-09-2020

Application type: First submission

Review commission: CMO regio Arnhem-Nijmegen (Nijmegen)

Study registrations

Followed up by the following (possibly more current) registration

No registrations found.

Other (possibly less up-to-date) registrations in this register

No registrations found.

In other registers

Register	ID
CCMO	NL74582.091.20