

# Protocol development for a Hyperfine Very Low Field (0.064 T) MRI system

Published: 02-09-2022

Last updated: 07-04-2024

Our overall aim is to measure several distinct MR parameters at 0.064 Tesla, and to use these to derive optimal MRI-protocols in terms of contrast and SNR. Some of these new developments may subsequently be used in clinical research protocols (which...

<b>Ethical review</b>	Approved WMO
<b>Status</b>	Recruiting
<b>Health condition type</b>	Other condition
<b>Study type</b>	Observational non invasive

## Summary

### ID

NL-OMON51553

### Source

ToetsingOnline

### Brief title

Very Low Field MRI protocol development

### Condition

- Other condition

### Synonym

N.A.

### Health condition

MR protocol development

### Research involving

Human

## Sponsors and support

**Primary sponsor:** Radiologie

**Source(s) of monetary or material Support:** NWO Domain - Applied and Engineering Sciences (AES)

## Intervention

**Keyword:** Brain, MRI, very low field

## Outcome measures

### Primary outcome

The main study parameters of this project are quantitative measurements of MR parameters. Specifically, we will measure T1, T2, T1\*, T2\*, magnetisation transfer and diffusion in the human brain. From this data, we will perform quantitative analysis (signal-to-noise ratio (SNR), contrast-to-noise ratio (CNR)). Having measured these, we will design optimized imaging protocols which will be quantitatively compared with those currently available on the system

### Secondary outcome

None

## Study description

### Background summary

Very low-field (<0.1 T) portable MRI scanners are emerging as a new category of systems with much lower costs than conventional systems and the ability to be used in sites and applications where conventional MRI is impossible. In spring 2022 the LUMC Department of Radiology, in collaboration with the Gates Foundation, obtained a Hyperfine Swoop 0.064 T portable MRI scanner. The purpose of our collaboration is to perform MRI-protocol development to optimize image contrast in the brain using a number of different MRI techniques. Since MR parameters are much less studied at very low field than at conventional clinical field strengths, this will involve measuring each MR parameter in a number of healthy volunteers to derive optimized parameters and show that image

contrast is improved compared to existing techniques on the system. Dissemination of findings will be performed via articles in world leading journals on MR techniques. Such developments and publication of results are common to the MR community, and lead to a continuous development and improvement of the capabilities of MR scanners.

## **Study objective**

Our overall aim is to measure several distinct MR parameters at 0.064 Tesla, and to use these to derive optimal MRI-protocols in terms of contrast and SNR. Some of these new developments may subsequently be used in clinical research protocols (which are not a part of this protocol), other developments are more fundamental technical MR developments for which applications will only benefit in the future.

## **Study design**

We note that there will be no data collection for medical research. Only projects aiming at the development, optimization and/or interpretation of non-invasive MR techniques are included. MRI-protocol development will follow the usual roadmap of MR physics research that consists of an iterative process of identification of new requirements or artefacts in existing techniques, MRI-protocol optimization, sequence development, pilot experiments, quality review meetings, and finally back to identification of sources of artefacts. For 6 different MR contrast mechanisms, we will scan 20 healthy volunteers (total number=120) to evaluate the consequent optimisation of Signal-to-Noise Ratio (SNR) and Contrast-to-Noise Ratio (CNR) and to publish results of these new protocols/sequences.

## **Study burden and risks**

All of the issues concerning safety and risk at 0.064 T are hundreds/thousands of times lower than the corresponding considerations at clinical MRI field strengths (1.5 and 3T), further details can be found in the risk and analysis sections of the attached IMDD. Although the Hyperfine Swoop currently does not have a CE mark, it was FDA approved in 2020 and so has been extensively tested with respect to safety and risk/benefit.

26-07-2022

## **Contacts**

### **Public**

Selecteer

Albinusdreef 2  
Leiden 2333 ZA  
NL  
**Scientific**  
Selecteer

Albinusdreef 2  
Leiden 2333 ZA  
NL

## **Trial sites**

### **Listed location countries**

Netherlands

## **Eligibility criteria**

### **Age**

Adults (18-64 years)

### **Inclusion criteria**

Healthy volunteers between the age >18 years or <65 years

### **Exclusion criteria**

- Age <18 years or >65 years
- Persons with reduced mental capacity
- Pregnancy and a chance of being pregnant (as reported by the volunteer)
- Not having a general practitioner
- MRI contraindication e.g. cardiac pacemaker, implants not approved for MRI (see [www.mrisafety.com](http://www.mrisafety.com)), claustrophobia, tinnitus, metal objects attached to the body that cannot be removed

## **Study design**

## Design

**Study type:** Observational non invasive

Masking: Open (masking not used)

Control: Uncontrolled

Primary purpose: Other

## Recruitment

NL

Recruitment status: Recruiting

Start date (anticipated): 01-03-2023

Enrollment: 120

Type: Actual

## Medical products/devices used

Generic name: Hyperfine Swoop

Registration: No

## Ethics review

Approved WMO

Date: 02-09-2022

Application type: First submission

Review commission: METC Leiden-Den Haag-Delft (Leiden)

metc-ldd@lumc.nl

## Study registrations

### Followed up by the following (possibly more current) registration

No registrations found.

## Other (possibly less up-to-date) registrations in this register

No registrations found.

## In other registers

### Register

CCMO

### ID

NL81853.058.22